SALVAFIX Standard



SALVAFIX standard perineal waist belt *with sewn perineal band and magnetic release safety locks. It can be installed before or after the patient lies down on the bed. The lower part, with two loop straps, has to be fixed to the bed (always on moving parts); the upper part has to be placed around the patient's waist. The lateral restraining bands must be always applied and used and they are intended to control or prevent the patient from rolling sideways, and let make the appropriated postural changes.*

Made of 100% polyester straps with inner reinforcement and stainless steel eyelets for a higher strength. 100% polyester soft to the touch piping for a comfortable fit. Safety locks are made of a Stainless Steel pin and a PC/ABS Casing with an internal mechanism that blocks the pin when it is inserted. Magnetic key is needed for system releasing, made of a PA casing and Neodymium magnet.

PERINEAL WAIST BELT

FEATURES



Adaptable to 80-100cm beds



Soft to the touch edges



Sewn perineal band



Sewn sidebands



Colour code for sizes
S (Yellow) - M (Blue) - L (Green)

SPECIFICATIONS

PRODUCT FAMILY: SALVAFIX Standard

PRODUCT NAME: Perineal waist belt

INTENDED USE: Intended to be used as preventing patient self-injury and falls from the bed or stretcher

BASIC UDI-DI: 843603535SVFNC

SIZE CHART

Ref.	Size	Measures (cm)	
		Contour Min./ Max.	Width
SVF3031P	S	47 / 76	16
SVF3030P	M	68 / 108	18
SVF3032P	L	95 / 135	18

COLOUR

- Natural white
- Light blue

COMPOSITION

Polyester
Stainless steel eyelets
LATEX FREE

LAUNDRY INSTRUCTIONS



- Machine wash 90ºC
- Do not bleach
- Tumble dry low temperature
- Do not iron

PACKAGING



- 1 - Salvafix Standard - Perineal waist belt
- 3 - Safety locks (4 safety locks for L - XL)
- 1 - Magnetic key
- 1 - Safety and instructions manual

MEDICARE SYSTEM

MEDI CARE SYSTEM S.L.U.
C/ Abat Escarré 15 - 08302 Mataró SPAIN
Tel.: +34 937 570 048 - Fax: +34 937 414 068
E-mail: info@medisystemsl.com · www.medicaresystem.es

CE Declaration according to Regulation (EU)745/2017 of the European Parliament & Council

*This data sheet could be changed without prior notice

UK CA CE